

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021086.D
 Acq On : 04 Dec 2021 00:04
 Operator : SY/VA
 Sample : VIBLK499
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK499

Quant Time: Dec 04 01:43:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 01:31:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	94895	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	92994	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	51354	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	38594	29.191	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 116.760%	
7) Chloroethane-d5	2.885	69	23297	26.322	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.280%	
11) 1,1-Dichloroethene-d2	4.019	63	41132	19.474	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 77.880%	
21) 2-Butanone-d5	7.080	46	15587	51.686	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 103.380%	
24) Chloroform-d	7.647	84	62503	26.300	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 105.200%	
26) 1,2-Dichloroethane-d4	8.305	65	35435	27.255	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 109.000%	
32) Benzene-d6	8.275	84	124826	26.252	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 105.000%	
36) 1,2-Dichloropropane-d6	9.274	67	34517	26.122	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 104.480%	
41) Toluene-d8	10.323	98	121802	24.982	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 99.920%	
43) trans-1,3-Dichloroprop...	10.579	79	15030	25.050	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 100.200%	
47) 2-Hexanone-d5	10.927	63	12514	49.499	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 99.000%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	28717	23.491	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 93.960%	
66) 1,2-Dichlorobenzene-d4	13.847	152	44628	24.254	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 97.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021086.D
Acq On : 04 Dec 2021 00:04
Operator : SY/VA
Sample : VIBLK499
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK499

Quant Time: Dec 04 01:43:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 01:31:24 2021
Response via : Initial Calibration

