

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021104.D
 Acq On : 04 Dec 2021 08:03
 Operator : SY/VA
 Sample : M4883-18
 Misc : 3.10g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9H4

Quant Time: Dec 04 09:31:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 04:50:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	4546	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	3471	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	904	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	2122	34.165	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 136.640%		
7) Chloroethane-d5	2.885	69	2744	64.848	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 259.400%#		
11) 1,1-Dichloroethene-d2	4.025	63	2784	28.020	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 112.080%#		
21) 2-Butanone-d5	7.092	46	1609	112.547	ug/L	0.02
Spiked Amount	50.000	Range 20 - 135	Recovery	= 225.100%#		
24) Chloroform-d	7.653	84	5263	46.879	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 187.520%#		
26) 1,2-Dichloroethane-d4	8.311	65	3346	55.645	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 222.600%#		
32) Benzene-d6	8.281	84	8471	48.690	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 194.760%#		
36) 1,2-Dichloropropane-d6	9.275	67	3101	64.319	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 257.280%#		
41) Toluene-d8	10.323	98	6549	36.974	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 147.880%#		
43) trans-1,3-Dichloroprop...	10.579	79	709	31.989	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 127.960%		
47) 2-Hexanone-d5	10.927	63	885	95.918	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 191.840%#		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	2492	56.109	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 224.440%#		
66) 1,2-Dichlorobenzene-d4	13.853	152	1005	31.690	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 126.760%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021104.D
Acq On : 04 Dec 2021 08:03
Operator : SY/VA
Sample : M4883-18
Misc : 3.10g/10.0mL/MSVOA_W/SOIL
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9H4

Quant Time: Dec 04 09:31:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 04:50:24 2021
Response via : Initial Calibration

