

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
 Data File : VW007187.D
 Acq On : 04 Dec 2018 15:04
 Operator : SY/AP
 Sample : VW1204SBL01
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1204SBL01

Quant Time: Dec 05 05:45:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 04 14:18:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	115872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	179395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	161271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	80538	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	53986	50.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.08%	
35) Dibromofluoromethane	7.88	113	50214	48.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.76%	
50) Toluene-d8	10.32	98	226882	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.62	95	81752	47.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
Data File : VW007187.D
Acq On : 04 Dec 2018 15:04
Operator : SY/AP
Sample : VW1204SBL01
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1204SBL01

Quant Time: Dec 05 05:45:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 04 14:18:19 2018
Response via : Initial Calibration

