

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
 Data File : VW007207.D
 Acq On : 05 Dec 2018 01:52
 Operator : SY/AP
 Sample : J6184-01RE
 Misc : 5.34G/5ML/MSVOA_W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 RO-331RE

Quant Time: Dec 05 07:44:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 04 14:18:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	94228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	147950	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	110929	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	35424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	42865	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	7.89	113	41840	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	10.32	98	158764	42.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.08%	
62) 4-Bromofluorobenzene	12.62	95	44500	31.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.18%	
Target Compounds						
16) Acetone	4.13	43	5382	25.914	ug/l #	89
17) Carbon Disulfide	4.38	76	7690	2.586	ug/l	98
52) Toluene	10.40	92	5771	2.102	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
Data File : VW007207.D
Acq On : 05 Dec 2018 01:52
Operator : SY/AP
Sample : J6184-01RE
Misc : 5.34G/5ML/MSVOA_W/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
RO-331RE

Quant Time: Dec 05 07:44:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 04 14:18:19 2018
Response via : Initial Calibration

