

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
 Data File : VW007219.D
 Acq On : 05 Dec 2018 06:59
 Operator : SY/AP
 Sample : J6210-05
 Misc : 6.31G/5ML/MSVOA_W/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WC-9(4.5)

Quant Time: Dec 05 08:07:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 04 14:18:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	80028	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	132367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	70007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	10787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	41025	55.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.12%	
35) Dibromofluoromethane	7.89	113	40301	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
50) Toluene-d8	10.33	98	124621	37.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.66%	
62) 4-Bromofluorobenzene	12.62	95	19271	15.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	30.58%	
Target Compounds						
16) Acetone	4.13	43	4560	25.852	ug/l	96
20) Methylene Chloride	4.92	84	14581	12.504	ug/l	91
21) trans-1,2-Dichloroethene	5.43	96	2314	2.705	ug/l	95
27) cis-1,2-Dichloroethene	7.17	96	50352	54.898	ug/l	90
44) Trichloroethene	9.09	130	207812	209.332	ug/l	98
64) Tetrachloroethene	10.87	164	764828	1495.319	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
Data File : VW007219.D
Acq On : 05 Dec 2018 06:59
Operator : SY/AP
Sample : J6210-05
Misc : 6.31G/5ML/MSVOA_W/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WC-9(4.5)

Quant Time: Dec 05 08:07:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 04 14:18:19 2018
Response via : Initial Calibration

