

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
 Data File : VW007221.D
 Acq On : 05 Dec 2018 07:50
 Operator : SY/AP
 Sample : J6208-03
 Misc : 6.03G/5ML/MSVOA_W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-1

Quant Time: Dec 05 11:12:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 05 01:35:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	65788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	104757	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	75230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	23062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	29281	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
35) Dibromofluoromethane	7.88	113	29366	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	10.33	98	103430	39.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.28%	
62) 4-Bromofluorobenzene	12.62	95	28857	28.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	57.86%	
Target Compounds						
16) Acetone	4.13	43	566	3.903	ug/l	97
20) Methylene Chloride	4.92	84	8216	6.739	ug/l	93
52) Toluene	10.39	92	5966	3.068	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120418\
Data File : VW007221.D
Acq On : 05 Dec 2018 07:50
Operator : SY/AP
Sample : J6208-03
Misc : 6.03G/5ML/MSVOA_W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-1

Quant Time: Dec 05 11:12:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 05 01:35:23 2018
Response via : Initial Calibration

