

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120518\
 Data File : VW007244.D
 Acq On : 05 Dec 2018 20:53
 Operator : SY/AP
 Sample : J6206-16RE
 Misc : 4.46G/5ML/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HLSP-SB-3A-(1)RE

Quant Time: Dec 06 06:59:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 05 01:35:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	101733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	160087	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	109842	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	25070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	42261	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
35) Dibromofluoromethane	7.88	113	40753	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
50) Toluene-d8	10.33	98	141533	35.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.10%	
62) 4-Bromofluorobenzene	12.62	95	32876	21.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	43.14%	
Target Compounds						
16) Acetone	4.14	43	1381	6.159	ug/l #	60
20) Methylene Chloride	4.93	84	11572	5.619	ug/l	96
52) Toluene	10.40	92	20208	6.801	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120518\
Data File : VW007244.D
Acq On : 05 Dec 2018 20:53
Operator : SY/AP
Sample : J6206-16RE
Misc : 4.46G/5ML/MSVOA_W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HLSP-SB-3A-(1)RE

Quant Time: Dec 06 06:59:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 05 01:35:23 2018
Response via : Initial Calibration

