

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW120518\
 Data File : VW007247.D
 Acq On : 05 Dec 2018 22:10
 Operator : SY/AP
 Sample : J6231-01
 Misc : 7.7G/5ML/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S49-0485

Quant Time: Dec 06 07:06:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 05 01:35:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	104858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	167347	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	148783	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	71396	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	38674	39.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.22%	
35) Dibromofluoromethane	7.89	113	36883	38.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.98%	
50) Toluene-d8	10.33	98	153366	36.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.66%	
62) 4-Bromofluorobenzene	12.62	95	55843	35.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.10%	
Target Compounds						
16) Acetone	4.13	43	8387	36.289	ug/l	Qvalue 100
17) Carbon Disulfide	4.38	76	6148	1.858	ug/l	98
20) Methylene Chloride	4.92	84	10618	4.362	ug/l #	93
27) cis-1,2-Dichloroethene	7.18	96	2977	2.477	ug/l	99
44) Trichloroethene	9.09	130	92797	73.937	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW120518\
Data File : VW007247.D
Acq On : 05 Dec 2018 22:10
Operator : SY/AP
Sample : J6231-01
Misc : 7.7G/5ML/MSVOA_W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S49-0485

Quant Time: Dec 06 07:06:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 05 01:35:23 2018
Response via : Initial Calibration

