

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW120618\
 Data File : VW007296.D
 Acq On : 06 Dec 2018 22:58
 Operator : SY/AP
 Sample : J6195-01
 Misc : 6.97G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-01-120518-A

Quant Time: Dec 07 02:22:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 05 01:35:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	104128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	164592	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	138073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	57960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	40568	41.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.70%	
35) Dibromofluoromethane	7.89	113	36966	39.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.44%	
50) Toluene-d8	10.33	98	139728	33.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.32%	
62) 4-Bromofluorobenzene	12.62	95	52543	33.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.06%	
Target Compounds						
					Qvalue	
16) Acetone	4.13	43	7651	33.337	ug/l	99
20) Methylene Chloride	4.93	84	9888	3.729	ug/l	91
52) Toluene	10.39	92	66758	21.853	ug/l	98
67) Ethyl Benzene	11.73	91	11594	2.063	ug/l	97
68) m/p-Xylenes	11.84	106	22790	10.336	ug/l	92
69) o-Xylene	12.17	106	8717	4.232	ug/l	89
84) 1,2,4-Trimethylbenzene	13.26	105	5113	1.242	ug/l	98
95) Naphthalene	15.38	128	5047	2.247	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW120618\
Data File : VW007296.D
Acq On : 06 Dec 2018 22:58
Operator : SY/AP
Sample : J6195-01
Misc : 6.97G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-01-120518-A

Quant Time: Dec 07 02:22:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W120418S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 05 01:35:23 2018
Response via : Initial Calibration

