

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
 Data File : VW021161.D
 Acq On : 06 Dec 2021 11:43
 Operator : SY/VA
 Sample : M4885-13RE
 Misc : 6.63g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9M0RE

Quant Time: Dec 07 01:49:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 07 01:47:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	45591	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	44721	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	25223	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	19157	30.159	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 120.640%		
7) Chloroethane-d5	2.885	69	13424	31.570	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 126.280%		
11) 1,1-Dichloroethene-d2	4.013	63	22727	22.397	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 89.600%		
21) 2-Butanone-d5	7.080	46	8151	56.258	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 112.520%		
24) Chloroform-d	7.653	84	30162	26.417	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 105.680%		
26) 1,2-Dichloroethane-d4	8.305	65	17939	28.719	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 114.880%		
32) Benzene-d6	8.275	84	60352	26.394	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 105.560%		
36) 1,2-Dichloropropane-d6	9.274	67	16711	26.298	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 105.200%		
41) Toluene-d8	10.323	98	56915	24.274	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 97.080%		
43) trans-1,3-Dichloroprop...	10.579	79	6929	24.014	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 96.040%		
47) 2-Hexanone-d5	10.920	63	6411	52.732	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 105.460%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	14588	24.815	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 99.240%		
66) 1,2-Dichlorobenzene-d4	13.853	152	20472	22.653	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 90.600%		
Target Compounds						
13) Acetone	4.135	43	690	5.228	ug/L	Qvalue 90
46) Tetrachloroethene	10.866	164	5723	8.908	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
Data File : VW021161.D
Acq On : 06 Dec 2021 11:43
Operator : SY/VA
Sample : M4885-13RE
Misc : 6.63g/10.0mL/MSVOA_W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9M0RE

Quant Time: Dec 07 01:49:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 07 01:47:32 2021
Response via : Initial Calibration

