

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
 Data File : VW021172.D
 Acq On : 06 Dec 2021 16:22
 Operator : SY/VA
 Sample : VIBLK507
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK507

Quant Time: Dec 07 01:51:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 07 01:47:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	96091	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	98166	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	58063	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	43494	32.487	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 129.960%	
7) Chloroethane-d5	2.879	69	28128	31.385	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 125.560%	
11) 1,1-Dichloroethene-d2	4.013	63	42929	20.072	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 80.280%	
21) 2-Butanone-d5	7.080	46	18604	60.922	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 121.840%	
24) Chloroform-d	7.647	84	68912	28.636	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 114.560%	
26) 1,2-Dichloroethane-d4	8.305	65	39798	30.230	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 120.920%	
32) Benzene-d6	8.275	84	138551	27.604	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 110.400%	
36) 1,2-Dichloropropane-d6	9.274	67	38086	27.305	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 109.200%	
41) Toluene-d8	10.323	98	135566	26.340	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 105.360%	
43) trans-1,3-Dichloroprop...	10.579	79	16675	26.328	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 105.320%	
47) 2-Hexanone-d5	10.920	63	14948	56.012	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 112.020%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	34168	26.478	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 105.920%	
66) 1,2-Dichlorobenzene-d4	13.853	152	53690	25.808	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 103.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
Data File : VW021172.D
Acq On : 06 Dec 2021 16:22
Operator : SY/VA
Sample : VIBLK507
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK507

Quant Time: Dec 07 01:51:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 07 01:47:32 2021
Response via : Initial Calibration

