

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
 Data File : VW021181.D
 Acq On : 06 Dec 2021 20:49
 Operator : SY/VA
 Sample : VW1206SBL02
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK419

Quant Time: Dec 07 07:23:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 07 07:23:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	110699	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	104506	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	55622	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	38916	25.232	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.920%	
7) Chloroethane-d5	2.879	69	26232	25.407	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 101.640%	
11) 1,1-Dichloroethene-d2	4.013	63	38669	15.694	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 62.760%	
21) 2-Butanone-d5	7.080	46	17927	50.958	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 101.920%	
24) Chloroform-d	7.647	84	64196	23.156	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 92.640%	
26) 1,2-Dichloroethane-d4	8.299	65	38107	25.126	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 100.520%	
32) Benzene-d6	8.275	84	128184	23.989	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 95.960%	
36) 1,2-Dichloropropane-d6	9.274	67	35962	24.218	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 96.880%	
41) Toluene-d8	10.323	98	127235	23.222	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 92.880%	
43) trans-1,3-Dichloroprop...	10.579	79	15024	22.282	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 89.120%	
47) 2-Hexanone-d5	10.920	63	14750	51.917	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 103.840%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	31991	23.287	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 93.160%	
66) 1,2-Dichlorobenzene-d4	13.853	152	48976	24.575	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 98.280%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120621\
Data File : VW021181.D
Acq On : 06 Dec 2021 20:49
Operator : SY/VA
Sample : VW1206SBL02
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK419

Quant Time: Dec 07 07:23:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 07 07:23:32 2021
Response via : Initial Calibration

