

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
 Data File : VW021197.D
 Acq On : 07 Dec 2021 10:59
 Operator : SY/VA
 Sample : M4868-05
 Misc : 5.04g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Dec 08 03:48:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 02:47:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	110983	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	104353	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	56428	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	40450	26.159	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.640%
7) Chloroethane-d5	2.885	69	29031	28.046	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.200%
11) 1,1-Dichloroethene-d2	4.019	63	46157	18.685	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.760%
21) 2-Butanone-d5	7.086	46	14253	40.411	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.820%
24) Chloroform-d	7.653	84	66369	23.879	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.520%
26) 1,2-Dichloroethane-d4	8.305	65	36569	24.050	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	8.275	84	133555	25.031	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.120%
36) 1,2-Dichloropropane-d6	9.275	67	36858	24.858	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.440%
41) Toluene-d8	10.323	98	131179	23.977	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.920%
43) trans-1,3-Dichloroprop...	10.579	79	14208	21.103	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.400%
47) 2-Hexanone-d5	10.927	63	11579	40.815	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.640%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	29222	21.302	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	50723	25.088	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
Data File : VW021197.D
Acq On : 07 Dec 2021 10:59
Operator : SY/VA
Sample : M4868-05
Misc : 5.04g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Dec 08 03:48:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 02:47:46 2021
Response via : Initial Calibration

