

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
 Data File : VW021199.D
 Acq On : 07 Dec 2021 13:05
 Operator : SY/VA
 Sample : M4888-04
 Misc : 3.85g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EX8B1

Quant Time: Dec 08 03:48:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 02:47:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	115768	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	106282	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	51251	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	46333	28.726	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	114.920%
7) Chloroethane-d5	2.885	69	29900	27.692	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.760%
11) 1,1-Dichloroethene-d2	4.019	63	49562	19.234	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.920%
21) 2-Butanone-d5	7.080	46	13838	37.613	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.220%
24) Chloroform-d	7.653	84	72549	25.023	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.080%
26) 1,2-Dichloroethane-d4	8.305	65	38059	23.995	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	8.275	84	143438	26.395	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.560%
36) 1,2-Dichloropropane-d6	9.275	67	39289	26.016	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.080%
41) Toluene-d8	10.323	98	138896	24.927	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.720%
43) trans-1,3-Dichloroprop...	10.579	79	15815	23.063	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.240%
47) 2-Hexanone-d5	10.921	63	10965	37.950	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.900%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	27247	19.502	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.000%
66) 1,2-Dichlorobenzene-d4	13.853	152	43420	23.645	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.600%
Target Compounds						
					Qvalue	
46) Tetrachloroethene	10.866	164	12574	8.235	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
Data File : VW021199.D
Acq On : 07 Dec 2021 13:05
Operator : SY/VA
Sample : M4888-04
Misc : 3.85g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EX8B1

Quant Time: Dec 08 03:48:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 02:47:46 2021
Response via : Initial Calibration

