

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
 Data File : VW021202.D
 Acq On : 07 Dec 2021 14:17
 Operator : SY/VA
 Sample : M4888-21
 Misc : 6.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EX8C6

Quant Time: Dec 08 03:49:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 02:47:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	107227	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	95086	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	42449	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	40862	27.352	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.400%
7) Chloroethane-d5	2.885	69	28049	28.047	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.200%
11) 1,1-Dichloroethene-d2	4.019	63	42402	17.766	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.080%
21) 2-Butanone-d5	7.080	46	15694	46.055	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.120%
24) Chloroform-d	7.653	84	64202	23.908	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.640%
26) 1,2-Dichloroethane-d4	8.305	65	37350	25.424	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.680%
32) Benzene-d6	8.275	84	126546	26.028	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.120%
36) 1,2-Dichloropropane-d6	9.274	67	35960	26.615	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.480%
41) Toluene-d8	10.323	98	118840	23.839	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.360%
43) trans-1,3-Dichloroprop...	10.579	79	14096	22.977	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.920%
47) 2-Hexanone-d5	10.921	63	12685	49.072	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.140%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	28719	22.976	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.920%
66) 1,2-Dichlorobenzene-d4	13.853	152	32392	21.297	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.200%
Target Compounds					Qvalue	
13) Acetone	4.123	43	1348	4.343	ug/L	94
34) Trichloroethene	9.092	95	2035	1.364	ug/L	94
46) Tetrachloroethene	10.860	164	28103	20.573	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
Data File : VW021202.D
Acq On : 07 Dec 2021 14:17
Operator : SY/VA
Sample : M4888-21
Misc : 6.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EX8C6

Quant Time: Dec 08 03:49:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 02:47:46 2021
Response via : Initial Calibration

