

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
 Data File : VW021208.D
 Acq On : 07 Dec 2021 16:41
 Operator : SY/VA
 Sample : VIBLK509
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK509

Quant Time: Dec 08 03:50:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 02:47:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	108469	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	100500	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	53696	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.342	65	39672	26.251	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.000%	
7) Chloroethane-d5	2.879	69	26791	26.482	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.920%	
11) 1,1-Dichloroethene-d2	4.013	63	40108	16.613	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 66.440%	
21) 2-Butanone-d5	7.092	46	10993	31.890	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	= 63.780%	
24) Chloroform-d	7.653	84	61284	22.560	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 90.240%	
26) 1,2-Dichloroethane-d4	8.305	65	32325	21.751	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 87.000%	
32) Benzene-d6	8.274	84	122809	23.899	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 95.600%	
36) 1,2-Dichloropropane-d6	9.274	67	33713	23.608	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 94.440%	
41) Toluene-d8	10.323	98	122312	23.213	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 92.840%	
43) trans-1,3-Dichloroprop...	10.579	79	12541	19.341	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 77.360%	
47) 2-Hexanone-d5	10.926	63	9455	34.606	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 69.220%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	25045	18.957	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 75.840%	
66) 1,2-Dichlorobenzene-d4	13.853	152	45667	23.736	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 94.960%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
Data File : VW021208.D
Acq On : 07 Dec 2021 16:41
Operator : SY/VA
Sample : VIBLK509
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK509

Quant Time: Dec 08 03:50:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 02:47:46 2021
Response via : Initial Calibration

