

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
 Data File : VW021210.D
 Acq On : 07 Dec 2021 17:29
 Operator : SY/VA
 Sample : VIBLK510
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK510

Quant Time: Dec 08 03:50:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 08 02:47:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	111488	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	102104	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	53730	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	34237	22.041	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.160%
7) Chloroethane-d5	2.885	69	24808	23.858	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.440%
11) 1,1-Dichloroethene-d2	4.013	63	35645	14.364	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.440%
21) 2-Butanone-d5	7.080	46	11575	32.670	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.340%
24) Chloroform-d	7.653	84	55764	19.972	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.880%
26) 1,2-Dichloroethane-d4	8.305	65	29059	19.024	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.080%
32) Benzene-d6	8.275	84	110806	21.224	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.880%
36) 1,2-Dichloropropane-d6	9.274	67	30255	20.854	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.400%
41) Toluene-d8	10.323	98	109240	20.407	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.640%
43) trans-1,3-Dichloroprop...	10.579	79	10906	16.555	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.240%
47) 2-Hexanone-d5	10.920	63	8917	32.124	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	22478	16.747	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	67.000%
66) 1,2-Dichlorobenzene-d4	13.853	152	41058	21.327	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120721\
Data File : VW021210.D
Acq On : 07 Dec 2021 17:29
Operator : SY/VA
Sample : VIBLK510
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK510

Quant Time: Dec 08 03:50:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 08 02:47:46 2021
Response via : Initial Calibration

