

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120921\
 Data File : VW021259.D
 Acq On : 09 Dec 2021 22:35
 Operator : SY/VA
 Sample : VIBLK523
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK523

Quant Time: Dec 10 00:34:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 10 00:28:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	92788	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	87826	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	45326	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42813	25.335	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 101.320%	
7) Chloroethane-d5	2.885	69	26430	23.658	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 94.640%	
11) 1,1-Dichloroethene-d2	4.019	63	40834	18.125	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 72.480%	
21) 2-Butanone-d5	7.080	46	12022	50.468	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 100.940%	
24) Chloroform-d	7.647	84	57876	22.382	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 89.520%	
26) 1,2-Dichloroethane-d4	8.305	65	31935	23.218	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 92.880%	
32) Benzene-d6	8.275	84	115751	24.044	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 96.160%	
36) 1,2-Dichloropropane-d6	9.274	67	31157	24.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 96.520%	
41) Toluene-d8	10.323	98	114836	23.347	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 93.400%	
43) trans-1,3-Dichloroprop...	10.579	79	12661	21.298	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 85.200%	
47) 2-Hexanone-d5	10.920	63	9672	50.661	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 101.320%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	24657	22.216	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 88.880%	
66) 1,2-Dichlorobenzene-d4	13.853	152	43941	25.014	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 100.040%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120921\
Data File : VW021259.D
Acq On : 09 Dec 2021 22:35
Operator : SY/VA
Sample : VIBLK523
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK523

Quant Time: Dec 10 00:34:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 10 00:28:10 2021
Response via : Initial Calibration

