

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
 Data File : VW021284.D
 Acq On : 10 Dec 2021 14:46
 Operator : SY/VA
 Sample : M5004-07
 Misc : 5.68g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGKW6

Quant Time: Dec 10 22:09:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 10 22:05:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	68754	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	57677	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	18734	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	28125	22.461	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.840%
7) Chloroethane-d5	2.885	69	22321	26.964	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.840%
11) 1,1-Dichloroethene-d2	4.013	63	30137	18.053	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.200%
21) 2-Butanone-d5	7.080	46	12438	70.467	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	140.940%#
24) Chloroform-d	7.647	84	47261	24.666	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.680%
26) 1,2-Dichloroethane-d4	8.305	65	25872	25.385	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.520%
32) Benzene-d6	8.275	84	87890	27.800	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	111.200%
36) 1,2-Dichloropropane-d6	9.274	67	25322	29.865	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	119.480%
41) Toluene-d8	10.323	98	79678	24.667	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.680%
43) trans-1,3-Dichloroprop...	10.579	79	10000	25.615	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	102.440%
47) 2-Hexanone-d5	10.927	63	8897	70.962	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	141.920%#
56) 1,1,2,2-Tetrachloroeth...	12.695	84	19692	27.017	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	16542	22.784	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.120%
Target Compounds					Qvalue	
13) Acetone	4.129	43	2244	14.181	ug/L	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
Data File : VW021284.D
Acq On : 10 Dec 2021 14:46
Operator : SY/VA
Sample : M5004-07
Misc : 5.68g/10.0mL/MSVOA_W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGKW6

Quant Time: Dec 10 22:09:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 10 22:05:53 2021
Response via : Initial Calibration

