

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
 Data File : VW021273.D
 Acq On : 10 Dec 2021 09:51
 Operator : SY/VA
 Sample : VW1210SBL01
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK420

Quant Time: Dec 10 22:07:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 10 22:05:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	90375	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	83783	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	45845	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	48073	29.207	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.840%
7) Chloroethane-d5	2.885	69	32416	29.791	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.160%
11) 1,1-Dichloroethene-d2	4.019	63	43498	19.822	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.280%
21) 2-Butanone-d5	7.098	46	11850	51.074	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.140%
24) Chloroform-d	7.647	84	57628	22.881	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.520%
26) 1,2-Dichloroethane-d4	8.311	65	32461	24.230	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.920%
32) Benzene-d6	8.275	84	113757	24.770	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.080%
36) 1,2-Dichloropropane-d6	9.274	67	31109	25.258	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.040%
41) Toluene-d8	10.323	98	114000	24.295	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.200%
43) trans-1,3-Dichloroprop...	10.579	79	13768	24.278	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.120%
47) 2-Hexanone-d5	10.927	63	9829	53.968	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.940%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	27552	26.023	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	43003	24.203	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
Data File : VW021273.D
Acq On : 10 Dec 2021 09:51
Operator : SY/VA
Sample : VW1210SBL01
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK420

Quant Time: Dec 10 22:07:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 10 22:05:53 2021
Response via : Initial Calibration

