

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
 Data File : VW021291.D
 Acq On : 10 Dec 2021 17:48
 Operator : SY/VA
 Sample : M5004-07RE
 Misc : 7.41g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGKW6RE

Quant Time: Dec 10 22:10:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 10 22:05:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	63200	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	54460	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	21181	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	24535	21.316	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.280%
7) Chloroethane-d5	2.885	69	18695	24.569	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.280%
11) 1,1-Dichloroethene-d2	4.013	63	23849	15.541	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.160%
21) 2-Butanone-d5	7.080	46	10174	62.705	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.420%
24) Chloroform-d	7.647	84	40442	22.962	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.840%
26) 1,2-Dichloroethane-d4	8.305	65	21879	23.354	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	8.275	84	71027	23.793	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.160%
36) 1,2-Dichloropropane-d6	9.275	67	21116	26.376	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.520%
41) Toluene-d8	10.323	98	63151	20.705	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	10.579	79	8361	22.682	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.720%
47) 2-Hexanone-d5	10.921	63	7197	60.794	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.580%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	17153	24.924	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	15807	19.256	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	77.040%
Target Compounds					Qvalue	
13) Acetone	4.123	43	2583	17.758	ug/L	59


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121021\
Data File : VW021291.D
Acq On : 10 Dec 2021 17:48
Operator : SY/VA
Sample : M5004-07RE
Misc : 7.41g/10.0mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGKW6RE

Quant Time: Dec 10 22:10:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 10 22:05:53 2021
Response via : Initial Calibration





#13
 Acetone
 Concen: 17.758 ug/L
 RT: 4.123 min Scan# 30
 Delta R.T. -0.037 min
 Lab File: VW021291.D
 Acq: 10 Dec 2021 17:48

Instrument :
 MSVOA_W
 ClientSampleId :
 BGKW6RE

Tgt Ion: 43 Resp: 2583
 Ion Ratio Lower Upper
 43 100
 58 5.6 0.0 52.8

