

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121218\
 Data File : VW007404.D
 Acq On : 11 Dec 2018 18:26
 Operator : SY/AP
 Sample : J6271-17
 Misc : 5.99G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F9E32

Quant Time: Dec 12 03:49:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 11 01:18:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	4261	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	3597	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	1138	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	811	29.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.88%
7) Chloroethane-d5	2.89	69	1162	49.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	197.76%#
10) 1,1-Dichloroethene-d2	4.02	63	2298	24.01	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	96.04%
20) 2-Butanone-d5	7.07	46	697	39.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.68%
24) Chloroform-d	7.64	84	1663	16.55	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	66.20%
26) 1,2-Dichloroethane-d4	8.31	65	2089	36.52	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	146.08%#
29) Benzene-d6	8.27	84	5939	30.40	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	121.60%
33) 1,2-Dichloropropane-d6	9.27	67	1781	31.95	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	127.80%#
37) Toluene-d8	10.32	98	5195	28.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	113.44%
38) trans-1,3-Dichloropropene-	10.57	79	752	26.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	105.56%
39) 2-Hexanone-d5	10.93	63	670	49.93	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.86%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	2112	41.58	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	166.32%#
61) 1,2-Dichlorobenzene-d4	13.86	152	1292	32.27	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	129.08%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121218\
Data File : VW007404.D
Acq On : 11 Dec 2018 18:26
Operator : SY/AP
Sample : J6271-17
Misc : 5.99G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F9E32

Quant Time: Dec 12 03:49:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120718S.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 11 01:18:15 2018
Response via : Initial Calibration

