

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
 Data File : VW021367.D
 Acq On : 13 Dec 2021 16:41
 Operator : SY/VA
 Sample : VIBLK533
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK533

Quant Time: Dec 14 00:37:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 14 00:34:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	79765	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	75049	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	41558	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	39856	27.435	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	109.760%	
7) Chloroethane-d5	2.879	69	27977	29.132	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	116.520%	
11) 1,1-Dichloroethene-d2	4.007	63	33906	17.507	ug/L	-0.01
Spiked Amount	25.000	Range 45 - 110	Recovery	=	70.040%	
21) 2-Butanone-d5	7.092	46	9757	47.647	ug/L	0.01
Spiked Amount	50.000	Range 20 - 135	Recovery	=	95.300%	
24) Chloroform-d	7.653	84	52131	23.452	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	93.800%	
26) 1,2-Dichloroethane-d4	8.305	65	26842	22.701	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	90.800%	
32) Benzene-d6	8.275	84	98906	24.043	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	96.160%	
36) 1,2-Dichloropropane-d6	9.274	67	27419	24.853	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	99.400%	
41) Toluene-d8	10.323	98	97550	23.209	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	92.840%	
43) trans-1,3-Dichloroprop...	10.579	79	12328	24.268	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	97.080%	
47) 2-Hexanone-d5	10.927	63	8121	49.779	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	99.560%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	23742	25.034	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	100.120%	
66) 1,2-Dichlorobenzene-d4	13.847	152	35889	22.283	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	89.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
 Data File : VW021367.D
 Acq On : 13 Dec 2021 16:41
 Operator : SY/VA
 Sample : VIBLK533
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 VIBLK533

Quant Time: Dec 14 00:37:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 14 00:34:04 2021
 Response via : Initial Calibration

