

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
 Data File : VW021357.D
 Acq On : 13 Dec 2021 12:38
 Operator : SY/VA
 Sample : M4885-23
 Misc : 5.05g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 14 00:35:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 14 00:34:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	90776	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	84175	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	45223	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42757	25.862	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 103.440%	
7) Chloroethane-d5	2.885	69	29585	27.069	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.280%	
11) 1,1-Dichloroethene-d2	4.019	63	43038	19.526	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 78.120%	
21) 2-Butanone-d5	7.086	46	11175	47.952	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 95.900%	
24) Chloroform-d	7.653	84	63209	24.987	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 99.960%	
26) 1,2-Dichloroethane-d4	8.305	65	32540	24.182	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 96.720%	
32) Benzene-d6	8.274	84	121812	26.401	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 105.600%	
36) 1,2-Dichloropropane-d6	9.274	67	33097	26.747	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 107.000%	
41) Toluene-d8	10.323	98	120335	25.526	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 102.120%	
43) trans-1,3-Dichloroprop...	10.579	79	14745	25.880	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 103.520%	
47) 2-Hexanone-d5	10.926	63	8810	48.148	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 96.300%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	26037	24.477	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 97.920%	
66) 1,2-Dichlorobenzene-d4	13.853	152	45678	26.062	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 104.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
Data File : VW021357.D
Acq On : 13 Dec 2021 12:38
Operator : SY/VA
Sample : M4885-23
Misc : 5.05g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 14 00:35:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 14 00:34:04 2021
Response via : Initial Calibration

