

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
 Data File : VW021360.D
 Acq On : 13 Dec 2021 13:50
 Operator : SY/VA
 Sample : M4888-23
 Misc : 5.02g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Dec 14 00:35:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Dec 14 00:34:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	88493	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	84266	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	45437	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	44076	27.348	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	= 109.400%	
7) Chloroethane-d5	2.885	69	29477	27.666	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 110.680%	
11) 1,1-Dichloroethene-d2	4.013	63	37691	17.541	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 70.160%	
21) 2-Butanone-d5	7.092	46	9724	42.802	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	= 85.600%	
24) Chloroform-d	7.647	84	52996	21.490	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 85.960%	
26) 1,2-Dichloroethane-d4	8.305	65	28613	21.812	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 87.240%	
32) Benzene-d6	8.275	84	106572	23.073	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 92.280%	
36) 1,2-Dichloropropane-d6	9.274	67	28679	23.152	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 92.600%	
41) Toluene-d8	10.323	98	101763	21.563	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 86.240%	
43) trans-1,3-Dichloroprop...	10.579	79	12586	22.066	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 88.280%	
47) 2-Hexanone-d5	10.927	63	7761	42.369	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 84.740%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	23752	22.305	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 89.240%	
66) 1,2-Dichlorobenzene-d4	13.853	152	38644	21.945	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 87.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121321\
Data File : VW021360.D
Acq On : 13 Dec 2021 13:50
Operator : SY/VA
Sample : M4888-23
Misc : 5.02g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Dec 14 00:35:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Dec 14 00:34:04 2021
Response via : Initial Calibration

