

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121324\
 Data File : VW031245.D
 Acq On : 13 Dec 2024 18:32
 Operator : SY/MD
 Sample : P5232-17
 Misc : 6.74g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28T1

Quant Time: Dec 14 00:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 14 00:44:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	801280	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	539499	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	155442	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	232003	19.243	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	76.960%	
7) Chloroethane-d5	2.899	69	170263	19.091	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	76.360%	
11) 1,1-Dichloroethene-d2	4.015	65	109105	19.078	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	76.320%	
21) 2-Butanone-d5	7.100	46	53524	23.574	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	47.140%	
24) Chloroform-d	7.648	84	397819	18.120	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	72.480%	
26) 1,2-Dichloroethane-d4	8.307	65	209668	19.016	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	76.080%	
32) Benzene-d6	8.270	84	799568	23.123	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	92.480%	
36) 1,2-Dichloropropane-d6	9.270	67	238190	23.635	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	94.520%	
41) Toluene-d8	10.319	98	651453	20.251	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	81.000%	
43) trans-1,3-Dichloroprop...	10.575	79	85949	18.496	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	74.000%	
47) 2-Hexanone-d5	10.922	63	46367	34.212	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	68.420%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	139117	21.986	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	87.960%	
66) 1,2-Dichlorobenzene-d4	13.848	152	105367	18.259	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	73.040%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121324\
Data File : VW031245.D
Acq On : 13 Dec 2024 18:32
Operator : SY/MD
Sample : P5232-17
Misc : 6.74g/10mL/MSVOA_W/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28T1

Quant Time: Dec 14 00:48:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 14 00:44:31 2024
Response via : Initial Calibration

