

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121324\
 Data File : VW031252.D
 Acq On : 13 Dec 2024 21:19
 Operator : SY/MD
 Sample : P5232-10RE
 Misc : 7.02g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28S4RE

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 12/16/2024
 Supervised By :Mahesh Dadoda 12/17/2024

Quant Time: Dec 14 00:49:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 14 00:44:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	469989	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	236922	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	52263	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.356	65	186353	26.353	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 105.400%			
7) Chloroethane-d5	2.887	69	128998	24.660	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 98.640%			
11) 1,1-Dichloroethene-d2	4.021	65	80154	23.895	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 95.600%			
21) 2-Butanone-d5	7.093	46	35821	26.898	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 53.800%			
24) Chloroform-d	7.648	84	256466	19.916	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 79.680%			
26) 1,2-Dichloroethane-d4	8.307	65	144924m	22.409	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 89.640%			
32) Benzene-d6	8.276	84	483398	31.833	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 127.320%			
36) 1,2-Dichloropropane-d6	9.270	67	149327	33.741	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 134.960%#			
41) Toluene-d8	10.318	98	347227	24.579	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 98.320%			
43) trans-1,3-Dichloroprop...	10.575	79	35970	17.627	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 70.520%			
47) 2-Hexanone-d5	10.922	63	21735	36.519	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 73.040%			
56) 1,1,2,2-Tetrachloroeth...	12.690	84	79708	28.685	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 114.720%			
66) 1,2-Dichlorobenzene-d4	13.848	152	36974	19.057	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 76.240%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121324\
Data File : VW031252.D
Acq On : 13 Dec 2024 21:19
Operator : SY/MD
Sample : P5232-10RE
Misc : 7.02g/10mL/MSVOA_W/SOIL/B
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28S4RE

Quant Time: Dec 14 00:49:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 14 00:44:31 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 12/16/2024
Supervised By :Mahesh Dadoda 12/17/2024

