

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121423\
 Data File : VW027631.D
 Acq On : 14 Dec 2023 16:33
 Operator : SY/MD
 Sample : 05799-18
 Misc : 3.62g/10mL/MSVOA_W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGRS8

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/15/2023
 Supervised By :Mahesh Dadoda 12/15/2023

Quant Time: Dec 15 03:33:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 15 03:28:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	633397	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	577980	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	283606	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	217060	17.029	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	68.120%		
7) Chloroethane-d5	2.905	69	165478	17.909	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	71.640%		
11) 1,1-Dichloroethene-d2	4.021	65	96843	17.956	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	71.840%		
21) 2-Butanone-d5	7.093	46	90911	27.219	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	54.440%		
24) Chloroform-d	7.654	84	382283	20.220	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.880%		
26) 1,2-Dichloroethane-d4	8.307	65	204817m	21.185	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	84.720%		
32) Benzene-d6	8.276	84	826765	20.310	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	81.240%		
36) 1,2-Dichloropropane-d6	9.276	67	254433	21.201	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	84.800%		
41) Toluene-d8	10.324	98	732077	20.554	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	82.200%		
43) trans-1,3-Dichloroprop...	10.581	79	92911	20.021	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	80.080%		
47) 2-Hexanone-d5	10.922	63	71188	31.837	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	63.680%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	164458	16.661	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	66.640%		
66) 1,2-Dichlorobenzene-d4	13.848	152	207889	20.912	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	83.640%		
Target Compounds						Qvalue
46) Tetrachloroethene	10.861	164	2394	0.308	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121423\
Data File : VW027631.D
Acq On : 14 Dec 2023 16:33
Operator : SY/MD
Sample : 05799-18
Misc : 3.62g/10mL/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
BGRS8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/15/2023
Supervised By :Mahesh Dadoda 12/15/2023


Quant Time: Dec 15 03:33:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120623SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 15 03:28:28 2023
Response via : Initial Calibration





#46
Tetrachloroethene
Concen: 0.308 ug/L
RT: 10.861 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VW027631.D
Acq: 14 Dec 2023 16:33

Instrument :
MSVOA_W
ClientSampleId :
BGRS8



Tgt Ion	Ratio	Lower	Upper
164	100		
129	96.9	71.0	131.8
131	69.2	61.2	113.6
166	113.1	86.6	160.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/15/2023
Supervised By :Mahesh Dadoda 12/15/2023

