

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121518\
 Data File : VW007450.D
 Acq On : 12 Dec 2018 21:03
 Operator : SY/AP
 Sample : VW1215SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1215SBL01

Quant Time: Dec 13 04:31:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	82198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	137773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	124149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	60692	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	45193	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
35) Dibromofluoromethane	7.88	113	41409	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
50) Toluene-d8	10.32	98	157015	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
62) 4-Bromofluorobenzene	12.62	95	58249	50.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121518\
Data File : VW007450.D
Acq On : 12 Dec 2018 21:03
Operator : SY/AP
Sample : VW1215SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1215SBL01

Quant Time: Dec 13 04:31:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

