

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121518\
 Data File : VW007469.D
 Acq On : 13 Dec 2018 05:15
 Operator : SY/AP
 Sample : J6329-01
 Misc : 5.04G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 071-PIPE-1

Manual Integrations
 APPROVED

apatel
 12/13/2018 4:40:42 PM

Quant Time: Dec 13 09:30:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	8385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	16844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	15463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	11459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	6376	68.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	137.08%	
35) Dibromofluoromethane	7.88	113	4997	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
50) Toluene-d8	10.32	98	18494	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	12.62	95	25392m	180.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	361.28%	
Target Compounds						
16) Acetone	4.13	43	1394	85.792	ug/l	89
20) Methylene Chloride	4.92	84	755	4.968	ug/l #	90
73) Isopropylbenzene	12.47	105	22989	26.182	ug/l	99
78) n-propylbenzene	12.80	91	14771	14.948	ug/l	87
85) sec-Butylbenzene	13.38	105	45053	50.768	ug/l #	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121518\
Data File : VW007469.D
Acq On : 13 Dec 2018 05:15
Operator : SY/AP
Sample : J6329-01
Misc : 5.04G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
071-PIPE-1

Manual Integrations
APPROVED

apatel
12/13/2018 4:40:42 PM

Quant Time: Dec 13 09:30:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

