

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121618\
 Data File : VW007488.D
 Acq On : 13 Dec 2018 15:10
 Operator : SY/AP
 Sample : J6319-09
 Misc : 4.24G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 063_SW-E-1

Quant Time: Dec 14 04:07:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	87796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	144925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	122284	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	50352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	36296	37.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.52%	
35) Dibromofluoromethane	7.88	113	34043	38.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.60%	
50) Toluene-d8	10.32	98	117370	35.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.36%	
62) 4-Bromofluorobenzene	12.62	95	42558	35.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.38%	
Target Compounds						
16) Acetone	4.13	43	1726	10.145	ug/l	98
68) m/p-Xylenes	11.84	106	1875	1.101	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121618\
Data File : VW007488.D
Acq On : 13 Dec 2018 15:10
Operator : SY/AP
Sample : J6319-09
Misc : 4.24G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
063_SW-E-1

Quant Time: Dec 14 04:07:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

