

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121618\
 Data File : VW007492.D
 Acq On : 13 Dec 2018 16:54
 Operator : SY/AP
 Sample : J6329-01RE
 Misc : 5.24G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 071-PIPE-1RE

Quant Time: Dec 14 04:17:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	82956	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	145770	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	105721	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	64960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	38408	41.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.46%	
35) Dibromofluoromethane	7.88	113	34848	39.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.98%	
50) Toluene-d8	10.32	98	138928	41.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.98%	
62) 4-Bromofluorobenzene	12.62	95	130914	107.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	215.24%	
Target Compounds						
16) Acetone	4.12	43	8321	51.762	ug/l #	84
73) Isopropylbenzene	12.47	105	96504	19.388	ug/l	99
78) n-propylbenzene	12.81	91	60744	10.844	ug/l	88
85) sec-Butylbenzene	13.39	105	210785	41.899	ug/l #	54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121618\
Data File : VW007492.D
Acq On : 13 Dec 2018 16:54
Operator : SY/AP
Sample : J6329-01RE
Misc : 5.24G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
071-PIPE-1RE

Quant Time: Dec 14 04:17:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

