

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121618\
 Data File : VW007493.D
 Acq On : 13 Dec 2018 17:20
 Operator : SY/AP
 Sample : J6329-02RE
 Misc : 4.86G/6ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 072-DUP-1RE

Manual Integrations
 APPROVED

MMDadoda
 12/14/2018 11:24:38 AM

Quant Time: Dec 14 15:52:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	82081	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	141380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	102727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	64318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	37096	40.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.48%	
35) Dibromofluoromethane	7.88	113	34622	40.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.90%	
50) Toluene-d8	10.32	98	161659	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
62) 4-Bromofluorobenzene	12.62	95	161532	136.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	273.82%	
Target Compounds						
16) Acetone	4.12	43	8072	50.749	ug/l #	88
73) Isopropylbenzene	12.47	105	136254	27.647	ug/l	100
78) n-propylbenzene	12.81	91	89499	16.137	ug/l	91
85) sec-Butylbenzene	13.39	105	232229	46.623	ug/l #	54
89) n-Butylbenzene	13.83	91	82287m	19.836	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121618\
Data File : VW007493.D
Acq On : 13 Dec 2018 17:20
Operator : SY/AP
Sample : J6329-02RE
Misc : 4.86G/6ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
072-DUP-1RE

Manual Integrations
APPROVED

MMDadoda
12/14/2018 11:24:38 AM

Quant Time: Dec 14 15:52:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

