

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW121618\
 Data File : VW007498.D
 Acq On : 13 Dec 2018 21:39
 Operator : SY/AP
 Sample : VW1216SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1216SBL02

Quant Time: Dec 14 04:36:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	87194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	147659	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	132800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	64990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	42345	43.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.54%	
35) Dibromofluoromethane	7.88	113	39437	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
50) Toluene-d8	10.32	98	144887	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
62) 4-Bromofluorobenzene	12.62	95	56371	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW121618\
Data File : VW007498.D
Acq On : 13 Dec 2018 21:39
Operator : SY/AP
Sample : VW1216SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1216SBL02

Quant Time: Dec 14 04:36:46 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

