

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW121618\
 Data File : VW007473.D
 Acq On : 13 Dec 2018 07:50
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 13 08:21:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.349	0.321	8.0	98	0.00
3 P	Chloromethane	0.293	0.273	6.8	94	0.00
4 C	Vinyl Chloride	0.343	0.341	0.6#	98	0.00
5 T	Bromomethane	0.231	0.242	-4.8	102	0.00
6 T	Chloroethane	0.189	0.204	-7.9	100	0.00
7 T	Trichlorofluoromethane	0.309	0.295	4.5	93	0.00
8 T	Diethyl Ether	0.242	0.251	-3.7	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.509	0.474	6.9	88	0.01
10 T	Methyl Iodide	0.844	0.858	-1.7	94	0.00
11 T	Tert butyl alcohol	0.038	0.037	2.6	96	0.00
12 CM	1,1-Dichloroethene	0.490	0.485	1.0#	93	0.00
13 T	Acrolein	0.025	0.026	-4.0	100	0.00
14 T	Allyl chloride	0.723	0.713	1.4	92	0.00
15 T	Acrylonitrile	0.102	0.106	-3.9	95	0.00
16 T	Acetone	0.097	0.093	4.1	86	0.00
17 T	Carbon Disulfide	1.487	1.441	3.1	92	0.00
18 T	Methyl Acetate	0.263	0.271	-3.0	95	0.00
19 T	Methyl tert-butyl Ether	0.740	0.800	-8.1	98	0.00
20 T	Methylene Chloride	0.615	0.541	12.0	92	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.544	0.9	92	0.00
22 T	Diisopropyl ether	1.390	1.441	-3.7	94	0.00
23 T	Vinyl Acetate	0.777	0.810	-4.2	93	0.00
24 P	1,1-Dichloroethane	0.939	0.942	-0.3	93	0.00
25 T	2-Butanone	0.126	0.131	-4.0	92	0.00
26 T	2,2-Dichloropropane	0.622	0.579	6.9	87	0.00
27 T	cis-1,2-Dichloroethene	0.601	0.612	-1.8	92	0.00
28 T	Bromochloromethane	0.353	0.379	-7.4	98	0.00
29 T	Tetrahydrofuran	0.080	0.083	-3.8	92	0.00
30 C	Chloroform	1.022	1.034	-1.2#	93	0.00
31 T	Cyclohexane	0.862	0.774	10.2	89	0.00
32 T	1,1,1-Trichloroethane	0.896	0.898	-0.2	94	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.587	-5.8	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.303	0.323	-6.6	93	0.00
36 T	1,1-Dichloropropene	0.456	0.455	0.2	92	0.00
37 T	Ethyl Acetate	0.170	0.168	1.2	93	0.00
38 T	Carbon Tetrachloride	0.496	0.497	-0.2	92	0.00
39 T	Methylcyclohexane	0.552	0.546	1.1	91	0.00
40 TM	Benzene	1.266	1.283	-1.3	91	0.00
41 T	Methacrylonitrile	0.097	0.119	-22.7#	106	0.00
42 TM	1,2-Dichloroethane	0.398	0.412	-3.5	93	0.00
43 T	Isopropyl Acetate	0.335	0.350	-4.5	93	0.00
44 TM	Trichloroethene	0.361	0.369	-2.2	94	0.00
45 C	1,2-Dichloropropane	0.293	0.307	-4.8#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW121618\
 Data File : VW007473.D
 Acq On : 13 Dec 2018 07:50
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 13 08:21:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.166	0.175	-5.4	97	0.00
47 T	Bromodichloromethane	0.436	0.451	-3.4	92	0.00
48 T	Methyl methacrylate	0.172	0.172	0.0	89	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	100	0.00
50 S	Toluene-d8	1.135	1.223	-7.8	94	0.00
51 T	4-Methyl-2-Pentanone	0.158	0.175	-10.8	97	0.00
52 CM	Toluene	0.819	0.862	-5.3#	94	0.00
53 T	t-1,3-Dichloropropene	0.424	0.448	-5.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.490	0.505	-3.1	92	0.00
55 T	1,1,2-Trichloroethane	0.226	0.242	-7.1	97	0.00
56 T	Ethyl methacrylate	0.272	0.307	-12.9	98	0.00
57 T	1,3-Dichloropropane	0.372	0.411	-10.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.145	-16.0	94	0.00
59 T	2-Hexanone	0.106	0.121	-14.2	102	0.00
60 T	Dibromochloromethane	0.286	0.312	-9.1	97	0.00
61 T	1,2-Dibromoethane	0.224	0.245	-9.4	102	0.00
62 S	4-Bromofluorobenzene	0.417	0.466	-11.8	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.360	0.357	0.8	100	0.00
65 PM	Chlorobenzene	1.042	1.065	-2.2	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.380	-3.5	100	0.00
67 C	Ethyl Benzene	1.792	1.854	-3.5#	98	0.00
68 T	m/p-Xylenes	0.697	0.718	-3.0	101	0.00
69 T	o-Xylene	0.651	0.688	-5.7	97	0.00
70 T	Styrene	1.051	1.140	-8.5	102	0.00
71 P	Bromoform	0.188	0.196	-4.3	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.831	3.845	-0.4	100	0.00
74 T	N-amyl acetate	0.696	0.735	-5.6	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.567	0.606	-6.9	98	0.00
76 T	1,2,3-Trichloropropane	0.412	0.458	-11.2	97	0.00
77 T	Bromobenzene	0.849	0.878	-3.4	98	0.00
78 T	n-propylbenzene	4.312	4.493	-4.2	99	0.00
79 T	2-Chlorotoluene	2.609	2.649	-1.5	96	0.00
80 T	1,3,5-Trimethylbenzene	3.219	3.326	-3.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.187	0.182	2.7	95	0.00
82 T	4-Chlorotoluene	2.750	2.767	-0.6	98	0.00
83 T	tert-Butylbenzene	2.837	2.937	-3.5	100	0.00
84 T	1,2,4-Trimethylbenzene	3.313	3.340	-0.8	99	0.00
85 T	sec-Butylbenzene	3.872	3.989	-3.0	105	0.00
86 T	p-Isopropyltoluene	3.479	3.566	-2.5	98	0.00
87 T	1,3-Dichlorobenzene	1.693	1.703	-0.6	98	0.00
88 T	1,4-Dichlorobenzene	1.702	1.706	-0.2	101	0.00
89 T	n-Butylbenzene	3.225	3.276	-1.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW121618\
Data File : VW007473.D
Acq On : 13 Dec 2018 07:50
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Dec 13 08:21:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.619	0.634	-2.4	101	0.00
91 T	1,2-Dichlorobenzene	1.517	1.570	-3.5	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.114	0.116	-1.8	97	0.00
93 T	1,2,4-Trichlorobenzene	1.044	1.036	0.8	99	0.00
94 T	Hexachlorobutadiene	0.554	0.530	4.3	95	0.00
95 T	Naphthalene	1.997	2.107	-5.5	98	0.00
96 T	1,2,3-Trichlorobenzene	0.907	0.912	-0.6	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6