

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121619\
 Data File : VW014372.D
 Acq On : 16 Dec 2019 16:41
 Operator : SY/VA
 Sample : K6277-06
 Misc : 5.05G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0C60

Quant Time: Dec 17 08:05:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120519S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 17 07:27:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	848611	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	623664	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	148640	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	281964	21.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.76%
7) Chloroethane-d5	2.89	69	245099	25.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.60%
10) 1,1-Dichloroethene-d2	4.01	63	388974	17.14	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.56%
20) 2-Butanone-d5	7.07	46	200158	59.77	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.54%
24) Chloroform-d	7.65	84	599635	27.75	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	111.00%
26) 1,2-Dichloroethane-d4	8.30	65	323922	28.59	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	114.36%
29) Benzene-d6	8.27	84	1212360	33.17	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	132.68%
33) 1,2-Dichloropropane-d6	9.27	67	405993	36.10	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	144.40%#
37) Toluene-d8	10.32	98	932072	28.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	112.80%
38) trans-1,3-Dichloropropene-	10.57	79	109388	23.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.76%
39) 2-Hexanone-d5	10.92	63	146891	65.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.38%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	311911	34.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	136.16%#
61) 1,2-Dichlorobenzene-d4	13.85	152	193600	34.09	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	136.36%#

Target Compounds					Ovalue
13) Acetone	4.12	43	11659	3.611 ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121619\
Data File : VW014372.D
Acq On : 16 Dec 2019 16:41
Operator : SY/VA
Sample : K6277-06
Misc : 5.05G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0C60

Quant Time: Dec 17 08:05:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM120519S.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 17 07:27:11 2019
Response via : Initial Calibration

