

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
 Data File : VW007527.D
 Acq On : 14 Dec 2018 11:41
 Operator : SY/AP
 Sample : J6322-10
 Misc : 7.73G/5ML/MSVOA_W/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-2A-G-2-120718

Quant Time: Dec 14 14:01:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	75612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	119660	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	70559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	52330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	35844	42.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.46%	
35) Dibromofluoromethane	7.88	113	33126	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
50) Toluene-d8	10.32	98	27789	10.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	20.46%	
62) 4-Bromofluorobenzene	12.62	95	40331	40.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.78%	
Target Compounds						
8) Diethyl Ether	3.68	74	761	2.083	ug/l	88
16) Acetone	4.13	43	3607	24.617	ug/l	94
20) Methylene Chloride	4.91	84	10706	10.246	ug/l	94
52) Toluene	10.39	92	4295	2.191	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
Data File : VW007527.D
Acq On : 14 Dec 2018 11:41
Operator : SY/AP
Sample : J6322-10
Misc : 7.73G/5ML/MSVOA_W/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-2A-G-2-120718

Quant Time: Dec 14 14:01:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

