

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
 Data File : VW007529.D
 Acq On : 14 Dec 2018 12:32
 Operator : SY/AP
 Sample : J6322-12
 Misc : 7.11G/5ML/MSVOA_W/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-2A-G-4-120718

Quant Time: Dec 14 14:05:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	72764	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	125423	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	95599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	31722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	36066	44.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.36%	
35) Dibromofluoromethane	7.88	113	33251	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
50) Toluene-d8	10.32	98	94450	33.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.36%	
62) 4-Bromofluorobenzene	12.62	95	31000	29.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.24%	
Target Compounds						
8) Diethyl Ether	3.68	74	793	2.255	ug/l #	53
20) Methylene Chloride	4.91	84	13165	14.274	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
Data File : VW007529.D
Acq On : 14 Dec 2018 12:32
Operator : SY/AP
Sample : J6322-12
Misc : 7.11G/5ML/MSVOA_W/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-2A-G-4-120718

Quant Time: Dec 14 14:05:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

