

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
 Data File : VW007554.D
 Acq On : 15 Dec 2018 01:30
 Operator : SY/AP
 Sample : J6391-13
 Misc : 3.82G/5ML/MSVOA_W/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS017-0612-01

Quant Time: Dec 15 04:04:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 15 03:35:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	55734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	88681	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	61213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	16780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	33512	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
35) Dibromofluoromethane	7.88	113	29337	54.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.30%	
50) Toluene-d8	10.32	98	92421	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
62) 4-Bromofluorobenzene	12.62	95	22249	30.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.12%	
Target Compounds						
16) Acetone	4.13	43	1652	15.296	ug/l	90
20) Methylene Chloride	4.92	84	3328	1.863	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
Data File : VW007554.D
Acq On : 15 Dec 2018 01:30
Operator : SY/AP
Sample : J6391-13
Misc : 3.82G/5ML/MSVOA_W/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS017-0612-01

Quant Time: Dec 15 04:04:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 15 03:35:28 2018
Response via : Initial Calibration

