

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
 Data File : VW007562.D
 Acq On : 15 Dec 2018 04:57
 Operator : SY/AP
 Sample : J6391-01
 Misc : 3.70G/5ML/MSVOA_W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS015-0612-01

Quant Time: Dec 15 06:03:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	37916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	66676	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	41338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	10362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	31362	74.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	149.12%	
35) Dibromofluoromethane	7.88	113	24846	61.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.12%	
50) Toluene-d8	10.32	98	63217	41.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.54%	
62) 4-Bromofluorobenzene	12.62	95	14782	26.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	53.14%	
Target Compounds						
16) Acetone	4.12	43	7971	108.486	ug/l	Qvalue 100
20) Methylene Chloride	4.90	84	3051	3.988	ug/l	87
40) Benzene	8.32	78	3239	1.919	ug/l #	89
52) Toluene	10.39	92	2335	2.138	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121718\
Data File : VW007562.D
Acq On : 15 Dec 2018 04:57
Operator : SY/AP
Sample : J6391-01
Misc : 3.70G/5ML/MSVOA_W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS015-0612-01

Quant Time: Dec 15 06:03:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

