

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW121718\
 Data File : VW007590.D
 Acq On : 15 Dec 2018 18:45
 Operator : SY/AP
 Sample : J6378-04
 Misc : 3.81G/5ML/MSVOA W/SOIL
 ALS Vial : 76 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SD009-0006-01

Quant Time: Dec 15 22:36:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	10758	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	20813	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	16013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	4766	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	13583	113.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	227.62%	
35) Dibromofluoromethane	7.88	113	9606	76.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	152.48%	
50) Toluene-d8	10.32	98	23276	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.62	95	6713	38.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.30%	
Target Compounds						
16) Acetone	4.12	43	11056	530.336	ug/l #	88
20) Methylene Chloride	4.91	84	1343	8.531	ug/l	87
25) 2-Butanone	7.17	43	38973	1431.988	ug/l	98
59) 2-Hexanone	10.97	43	16818	382.678	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW121718\
Data File : VW007590.D
Acq On : 15 Dec 2018 18:45
Operator : SY/AP
Sample : J6378-04
Misc : 3.81G/5ML/MSVOA W/SOIL
ALS Vial : 76 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SD009-0006-01

Quant Time: Dec 15 22:36:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

