

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
 Data File : VW021421.D
 Acq On : 17 Dec 2021 09:58
 Operator : SY/VA
 Sample : VW1217SBL01
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK420

Quant Time: Dec 17 15:15:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 17 00:58:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.835	114	93271	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	93138	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	53291	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	40817	24.028	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.120%
7) Chloroethane-d5	2.879	69	27602	24.579	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.320%
11) 1,1-Dichloroethene-d2	4.013	63	39078	17.255	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.040%
21) 2-Butanone-d5	7.074	46	16116	67.304	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.600%
24) Chloroform-d	7.647	84	64735	24.905	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.640%
26) 1,2-Dichloroethane-d4	8.299	65	36783	26.604	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.400%
32) Benzene-d6	8.269	84	123461	24.183	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.720%
36) 1,2-Dichloropropane-d6	9.268	67	35092	25.630	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.520%
41) Toluene-d8	10.323	98	124021	23.776	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.120%
43) trans-1,3-Dichloroprop...	10.579	79	14648	23.235	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.960%
47) 2-Hexanone-d5	10.920	63	13111	64.758	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.520%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	32112	27.283	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.120%
66) 1,2-Dichlorobenzene-d4	13.847	152	48333	23.402	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
Data File : VW021421.D
Acq On : 17 Dec 2021 09:58
Operator : SY/VA
Sample : VW1217SBL01
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK420

Quant Time: Dec 17 15:15:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 17 00:58:26 2021
Response via : Initial Calibration

