

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
 Data File : VW021427.D
 Acq On : 17 Dec 2021 13:33
 Operator : SY/VA
 Sample : VIBLK543
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK543

Quant Time: Dec 17 21:16:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 17 00:58:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	84274	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	84614	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	49190	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	38636	25.173	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.680%	
7) Chloroethane-d5	2.879	69	26495	26.112	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 104.440%	
11) 1,1-Dichloroethene-d2	4.007	63	34980	17.095	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 68.360%	
21) 2-Butanone-d5	7.074	46	15047	69.548	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 139.100%#	
24) Chloroform-d	7.647	84	55260	23.530	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 94.120%	
26) 1,2-Dichloroethane-d4	8.299	65	32187	25.765	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 103.080%	
32) Benzene-d6	8.269	84	110197	23.760	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 95.040%	
36) 1,2-Dichloropropane-d6	9.268	67	30310	24.368	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 97.480%	
41) Toluene-d8	10.323	98	108389	22.873	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 91.480%	
43) trans-1,3-Dichloroprop...	10.573	79	12576	21.958	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 87.840%	
47) 2-Hexanone-d5	10.921	63	11659	63.388	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 126.780%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	29001	27.122	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 108.480%	
66) 1,2-Dichlorobenzene-d4	13.847	152	43466	22.800	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 91.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
Data File : VW021427.D
Acq On : 17 Dec 2021 13:33
Operator : SY/VA
Sample : VIBLK543
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK543

Quant Time: Dec 17 21:16:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 17 00:58:26 2021
Response via : Initial Calibration

