

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
 Data File : VW021438.D
 Acq On : 17 Dec 2021 18:07
 Operator : SY/VA
 Sample : VIBLK544
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK544

Quant Time: Dec 17 21:18:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Dec 17 00:58:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	125562	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	128055	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	74198	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	39898	17.447	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.800%
7) Chloroethane-d5	2.873	69	25239	16.695	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.800%
11) 1,1-Dichloroethene-d2	4.007	63	44696	14.660	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.640%
21) 2-Butanone-d5	7.074	46	18888	58.595	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.180%
24) Chloroform-d	7.647	84	69413	19.837	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.360%
26) 1,2-Dichloroethane-d4	8.305	65	41964	22.546	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	8.269	84	138044	19.667	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.680%
36) 1,2-Dichloropropane-d6	9.274	67	39055	20.747	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.000%
41) Toluene-d8	10.323	98	138630	19.330	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.320%
43) trans-1,3-Dichloroprop...	10.573	79	16677	19.240	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.960%
47) 2-Hexanone-d5	10.921	63	16382	58.851	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.700%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	37338	23.073	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	57505	19.998	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121721\
Data File : VW021438.D
Acq On : 17 Dec 2021 18:07
Operator : SY/VA
Sample : VIBLK544
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK544

Quant Time: Dec 17 21:18:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Dec 17 00:58:26 2021
Response via : Initial Calibration

