

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121724\
 Data File : VW031313.D
 Acq On : 17 Dec 2024 13:25
 Operator : SY/MD
 Sample : P5258-08RE
 Misc : 4.24g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E28X2RE

Quant Time: Dec 18 00:17:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 18 00:14:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	1456449	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	1211654	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	468804	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	325219	14.841	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.360%
7) Chloroethane-d5	2.887	69	266263	16.425	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.720%
11) 1,1-Dichloroethene-d2	4.015	65	154633	14.876	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.520%
21) 2-Butanone-d5	7.094	46	159300	38.600	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.200%
24) Chloroform-d	7.648	84	668087	16.742	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	66.960%
26) 1,2-Dichloroethane-d4	8.301	65	387255	19.323	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.280%
32) Benzene-d6	8.276	84	1336806	17.214	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.840%
36) 1,2-Dichloropropane-d6	9.270	67	406089	17.942	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.760%
41) Toluene-d8	10.319	98	1171697	16.218	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	64.880%
43) trans-1,3-Dichloroprop...	10.575	79	169597	16.251	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.000%
47) 2-Hexanone-d5	10.922	63	122958	40.396	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.800%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	307528	21.640	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.560%
66) 1,2-Dichlorobenzene-d4	13.849	152	308431	17.722	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	70.880%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121724\
Data File : VW031313.D
Acq On : 17 Dec 2024 13:25
Operator : SY/MD
Sample : P5258-08RE
Misc : 4.24g/10mL/MSVOA_W/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E28X2RE

Quant Time: Dec 18 00:17:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM121324SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 18 00:14:15 2024
Response via : Initial Calibration

