

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121818\
 Data File : VW007616.D
 Acq On : 16 Dec 2018 07:42
 Operator : SY/AP
 Sample : J6378-22
 Misc : 4.97G/5ML/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS002-0006-01

Quant Time: Dec 17 00:18:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	1498	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	1328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	260	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	1383	210.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	420.44%	
35) Dibromofluoromethane	7.88	113	913	100.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	201.36%	
50) Toluene-d8	10.32	98	1466	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
62) 4-Bromofluorobenzene	12.62	95	407	32.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.12%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.92	84	48	4.037	ug/l #	10
52) Toluene	10.39	92	1185	48.284	ug/l	93
67) Ethyl Benzene	11.73	91	266	5.590	ug/l #	48
68) m/p-Xylenes	11.83	106	933	50.431	ug/l	88
69) o-Xylene	12.17	106	35	2.026	ug/l #	1
76) 1,2,3-Trichloropropane	12.63	75	231	107.754	ug/l #	1
81) trans-1,4-Dichloro-2-buten	12.63	75	231	238.029	ug/l #	7
95) Naphthalene	15.36	128	91	8.765	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121818\
Data File : VW007616.D
Acq On : 16 Dec 2018 07:42
Operator : SY/AP
Sample : J6378-22
Misc : 4.97G/5ML/MSVOA_W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS002-0006-01

Quant Time: Dec 17 00:18:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

