

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW121818\
 Data File : VW007618.D
 Acq On : 16 Dec 2018 08:34
 Operator : SY/AP
 Sample : J6378-24
 Misc : 6.62G/5ML/MSVOA_W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P001-SS002-1218-01

Quant Time: Dec 17 00:18:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 11 02:58:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	2080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	3960	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	2749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	695	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.30	65	1675	72.59	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	145.18%	
35) Dibromofluoromethane	7.89	113	1539	64.20	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	128.40%	
50) Toluene-d8	10.33	98	3945	43.89	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	87.78%	
62) 4-Bromofluorobenzene	12.62	95	1118	33.83	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	67.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
36) 1,1-Dichloropropene	7.95	75	119	3.294	ug/l #	40
38) Carbon Tetrachloride	7.95	117	111	2.824	ug/l #	13
52) Toluene	10.38	92	1018	15.691	ug/l	91
67) Ethyl Benzene	11.74	91	135	1.370	ug/l #	48
68) m/p-Xylenes	11.84	106	673	17.573	ug/l	89
76) 1,2,3-Trichloropropane	12.62	75	587	102.435	ug/l #	1
81) trans-1,4-Dichloro-2-buten	12.62	75	587	226.279	ug/l #	7

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW121818\
Data File : VW007618.D
Acq On : 16 Dec 2018 08:34
Operator : SY/AP
Sample : J6378-24
Misc : 6.62G/5ML/MSVOA_W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P001-SS002-1218-01

Quant Time: Dec 17 00:18:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 11 02:58:16 2018
Response via : Initial Calibration

