


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121818\
 Data File : VW007597.D
 Acq On : 15 Dec 2018 22:38
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W121118S.M
 Title : SW846 8260
 Last Update : Tue Dec 11 02:58:16 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	9557	PASS
75	95	30	60	51.8	27514	PASS
95	95	100	100	100.0	53157	PASS
96	95	5	9	6.6	3494	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.9	39810	PASS
175	174	5	9	8.2	3258	PASS
176	174	95	101	97.5	38824	PASS
177	176	5	9	6.9	2682	PASS