

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\Vw121820\
 Data File : Vw017666.D
 Acq On : 18 Dec 2020 10:46
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 00:08:31 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W121420S.M
 Quant Title : Sw846 8260
 QLast Update : Mon Dec 14 12:39:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.392	0.399	-1.8	93	0.00
3 P	Chloromethane	0.442	0.361	18.3	78	0.00
4 C	Vinyl Chloride	0.640	0.622	2.8#	88	0.00
5 T	Bromomethane	0.492	0.525	-6.7	94	0.00
6 T	Chloroethane	0.400	0.423	-5.7	94	0.00
7 T	Trichlorofluoromethane	0.507	0.573	-13.0	99	0.00
8 T	Diethyl Ether	0.278	0.238	14.4	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.550	0.528	4.0	85	0.00
10 T	Methyl Iodide	0.810	0.763	5.8	85	0.00
11 T	Tert butyl alcohol	0.056	0.030	46.4#	57	0.00
12 CM	1,1-Dichloroethene	0.565	0.531	6.0#	85	0.00
13 T	Acrolein	0.037	0.032	13.5	75	0.00
14 T	Allyl chloride	0.870	0.828	4.8	85	0.00
15 T	Acrylonitrile	0.123	0.098	20.3	67	0.00
16 T	Acetone	0.133	0.135	-1.5	78	0.00
17 T	Carbon Disulfide	1.651	1.543	6.5	82	0.00
18 T	Methyl Acetate	0.269	0.214	20.4	69	0.00
19 T	Methyl tert-butyl Ether	0.764	0.712	6.8	80	0.00
20 T	Methylene Chloride	0.665	0.565	15.0	85	0.00
21 T	trans-1,2-Dichloroethene	0.627	0.605	3.5	86	0.00
22 T	Diisopropyl ether	1.587	1.505	5.2	84	-0.01
23 T	Vinyl Acetate	1.068	0.929	13.0	75	0.00
24 P	1,1-Dichloroethane	1.077	1.058	1.8	88	0.00
25 T	2-Butanone	0.176	0.139	21.0	66	0.00
26 T	2,2-Dichloropropane	0.792	0.719	9.2	88	0.00
27 T	cis-1,2-Dichloroethene	0.670	0.651	2.8	87	0.00
28 T	Bromochloromethane	0.458	0.462	-0.9	88	0.00
29 T	Tetrahydrofuran	0.099	0.076	23.2	65	0.00
30 C	Chloroform	1.094	1.097	-0.3#	89	0.00
31 T	Cyclohexane	1.052	0.967	8.1	85	0.00
32 T	1,1,1-Trichloroethane	0.935	0.938	-0.3	87	0.00
33 S	1,2-Dichloroethane-d4	0.609	0.558	8.4	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.331	0.317	4.2	82	0.00
36 T	1,1-Dichloropropene	0.523	0.535	-2.3	87	0.00
37 T	Ethyl Acetate	0.219	0.188	14.2	70	0.00
38 T	Carbon Tetrachloride	0.526	0.537	-2.1	85	0.00
39 T	Methylcyclohexane	0.663	0.667	-0.6	83	0.00
40 TM	Benzene	1.387	1.433	-3.3	88	0.00
41 T	Methacrylonitrile	0.124	0.099	20.2	63	0.00
42 TM	1,2-Dichloroethane	0.422	0.429	-1.7	85	0.00
43 T	Isopropyl Acetate	0.410	0.349	14.9	69	0.00
44 TM	Trichloroethene	0.383	0.387	-1.0	85	0.00
45 C	1,2-Dichloropropane	0.342	0.344	-0.6#	85	0.00
46 T	Dibromomethane	0.191	0.188	1.6	83	0.00
47 T	Bromodichloromethane	0.488	0.497	-1.8	86	0.00
48 T	Methyl methacrylate	0.201	0.171	14.9	68	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	61	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\Vw121820\
 Data File : Vw017666.D
 Acq On : 18 Dec 2020 10:46
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 00:08:31 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W121420S.M
 Quant Title : Sw846 8260
 QLast Update : Mon Dec 14 12:39:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 S	Toluene-d8	1.288	1.246	3.3	82	0.00
51 T	4-Methyl-2-Pentanone	0.204	0.166	18.6	66	0.00
52 CM	Toluene	0.907	0.931	-2.6#	88	0.00
53 T	t-1,3-Dichloropropene	0.499	0.487	2.4	82	0.00
54 T	cis-1,3-Dichloropropene	0.577	0.570	1.2	83	0.00
55 T	1,1,2-Trichloroethane	0.256	0.249	2.7	82	0.00
56 T	Ethyl methacrylate	0.331	0.290	12.4	71	0.00
57 T	1,3-Dichloropropane	0.452	0.435	3.8	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.158	0.153	3.2	84	0.00
59 T	2-Hexanone	0.137	0.123	10.2	69	0.00
60 T	Dibromochloromethane	0.320	0.310	3.1	80	0.00
61 T	1,2-Dibromoethane	0.253	0.233	7.9	77	0.00
62 S	4-Bromofluorobenzene	0.460	0.438	4.8	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.341	0.350	-2.6	88	0.00
65 PM	Chlorobenzene	1.075	1.100	-2.3	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.386	-1.3	84	0.00
67 C	Ethyl Benzene	1.997	2.108	-5.6#	89	0.00
68 T	m/p-Xylenes	0.762	0.799	-4.9	89	0.00
69 T	o-Xylene	0.718	0.734	-2.2	87	0.00
70 T	Styrene	1.200	1.237	-3.1	87	0.00
71 P	Bromoform	0.204	0.180	11.8	71	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.954	4.161	-5.2	89	0.00
74 T	N-aryl acetate	0.829	0.701	15.4	69	0.00
75 P	1,1,2,2-Tetrachloroethane	0.659	0.597	9.4	74	0.00
76 T	1,2,3-Trichloropropane	0.481	0.493	-2.5	87	0.00
77 T	Bromobenzene	0.847	0.847	0.0	84	0.00
78 T	n-propylbenzene	4.694	5.027	-7.1	89	0.00
79 T	2-Chlorotoluene	2.617	2.786	-6.5	90	0.00
80 T	1,3,5-Trimethylbenzene	3.346	3.498	-4.5	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.238	0.193	18.9	66	0.00
82 T	4-Chlorotoluene	2.762	2.916	-5.6	90	0.00
83 T	tert-Butylbenzene	2.862	3.024	-5.7	87	0.00
84 T	1,2,4-Trimethylbenzene	3.296	3.484	-5.7	89	0.00
85 T	sec-Butylbenzene	4.019	4.358	-8.4	90	0.00
86 T	p-Isopropyltoluene	3.628	3.919	-8.0	91	0.00
87 T	1,3-Dichlorobenzene	1.688	1.792	-6.2	88	0.00
88 T	1,4-Dichlorobenzene	1.675	1.742	-4.0	88	0.00
89 T	n-Butylbenzene	3.565	3.878	-8.8	91	0.00
90 T	Hexachloroethane	0.713	0.752	-5.5	88	0.00
91 T	1,2-Dichlorobenzene	1.500	1.496	0.3	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.100	16.7	66	0.00
93 T	1,2,4-Trichlorobenzene	0.960	0.968	-0.8	84	0.00
94 T	Hexachlorobutadiene	0.504	0.528	-4.8	87	0.00
95 T	Naphthalene	1.963	1.748	11.0	74	0.00
96 T	1,2,3-Trichlorobenzene	0.830	0.801	3.5	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_w\Data\Vw121820\
Data File : Vw017666.D
Acq On : 18 Dec 2020 10:46
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_w/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Dec 19 00:08:31 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_w\Method\82W121420S.M
Quant Title : Sw846 8260
QLast Update : Mon Dec 14 12:39:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)
----------	-------	------	------------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6