

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121821\
 Data File : VW021472.D
 Acq On : 18 Dec 2021 13:29
 Operator : SY/VA
 Sample : M5154-16
 Misc : 7.05g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGL96

Quant Time: Dec 18 23:11:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 18 00:55:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	188996	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	183471	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	94954	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.348	65	51483	14.957	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.840%
7) Chloroethane-d5	2.885	69	29999	13.183	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.720%
11) 1,1-Dichloroethene-d2	4.013	63	70003	15.255	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.000%
21) 2-Butanone-d5	7.073	46	24903	51.325	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.660%
24) Chloroform-d	7.646	84	103587	19.668	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.680%
26) 1,2-Dichloroethane-d4	8.305	65	57887	20.662	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.640%
32) Benzene-d6	8.274	84	213470	21.227	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.920%
36) 1,2-Dichloropropane-d6	9.274	67	58348	21.634	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	86.520%
41) Toluene-d8	10.323	98	209192	20.359	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.440%
43) trans-1,3-Dichloroprop...	10.579	79	24577	19.790	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.160%
47) 2-Hexanone-d5	10.920	63	21773	54.593	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.180%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	48113	20.752	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.000%
66) 1,2-Dichlorobenzene-d4	13.852	152	71933	19.547	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.200%
Target Compounds						
					Qvalue	
13) Acetone	4.123	43	10147	23.328	ug/L	83
22) 2-Butanone	7.177	43	3612	6.359	ug/L	85
29) Cyclohexane	7.921	56	4342	1.189	ug/L	# 50
53) m,p-Xylene	11.835	106	4523	0.839	ug/L	75
62) 1,3,5-Trimethylbenzene	12.938	105	11671	1.063	ug/L	86
63) 1,2,4-Trimethylbenzene	13.249	105	31327	2.892	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW121821\
Data File : VW021472.D
Acq On : 18 Dec 2021 13:29
Operator : SY/VA
Sample : M5154-16
Misc : 7.05g/10.0mL/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGL96

Quant Time: Dec 18 23:11:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 18 00:55:20 2021
Response via : Initial Calibration

